

TECHNICAL DATA SHEET

Perfringens Agar Base (O.P.S.P.)

Principle

Perfringens Agar (O.P.S.P.) is based on the formula developed by Handford (1974) and media is composed of tryptone, soya peptone, yeast extract, liver extract, ferric ammonium citrate, sodium metabisulphite, tris base and agar. Tryptone, soya peptone, liver extract and yeast extract serve as source of carbon, nitrogen, amino acids and other necessary elements for the growth. Sodium metabisulphite and ferric ammonium citrate is indicator for hydrogen sulphate production or sulphate reduction. The tris provide buffering capacity. The media is fortified with antibiotics like polymyxin B, sulphadiazine and oleandomycin to increase selective of medium and to inhibit sulphite reducing bacteria other than *C. perfringenes*.

Use: For presumptive identification of *Clostridium perfringens* in foods

Contents*

Ingredients	Gram/Litre
Tryptone	15.000
Soya peptone	5.000
Yeast extract	5.000
Liver extract	7.000
Ferric ammonium citrate	1.000
Sodium metabisulphite	1.000
Tris	1.500
Agar	15.000
pH at 25°C	7.3 ±0.2

* Formula adjusted for optimum performance and parameters

Directions: Dissolve 50.50 grams in 1000 ml distilled water. Boil to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121 °C) for 15 min, cool it to 42-45 °C and add 100 mg of sodium sulphadiazine and 10000 IU of polymyxin B sulphate and 0.5 mg of oleandomycin phosphate. Mix well and distribute aseptically desired and inoculate test sample aseptically.

OXFORD LAB FINE CHEM LLP

ISO 9001-2008 Certified Company

Regd Office: Unit no 12, 1st Floor,
Neminath Industrial Estate No.6,
Navghar, Vasai (East), Palghar - 410210.
Maharashtra, INDIA.

Tel: +91 250 2390032 / 2390989 / 2390990
Email: sales@oxfordlabchem.com /
info@oxfordlabchem.com
Web: www.oxfordlabchem.com



Specimens types analyzed

Food samples and dairy products.

Precautions to be taken

These microbial media are intended for the in-vitro use only. All the handling, experiments, storage, and discarding should be performed with the help of skilled and knowledgeable technicians and as per the established guidelines. The material should be disposed only after proper sterilization by autoclaving. Please go through the MSDS of the media to avoid any accidents or in emergency.

Performance and Evaluation

The expected performance of the medium is liable to use as per the direction on the label when stored at optimum conditions and within expiry date.

Quality Control

Appearance	Light beige colored, free-flowing, homogeneous
Reaction of 5.05% solution	7.30 ±0.2 at 25 °C
pH	7.10- 7.50
Gelling strength	1.5% agar
Color and clarity of ready medium	Light amber colored, clear opalescent gel
Growth Promotion properties	Best at ≤ 100 CFU at 32-37 °C for 18-72 h
Indicative properties	Optimum at ≤ 100 CFU at 32-37 °C for 18-48 h
Negative control	Performed using sterile distilled water

Different Microbial Response: Cultural characteristics observed after an incubation anaerobically at 35±2°C for 18-48 hours.

Organism	ATCC	Growth	Recovery	Color of colony
<i>Clostridium perfringens</i>	12924	Luxuriant	≥ 60%	Black
<i>Escherichia coli</i>	8739	Inhibited	--	--
<i>Bacillus subtilis</i>	6633	Inhibited	--	--
<i>Staphylococcus aureus</i>	25923	Inhibited	--	--

This document has been produced electronically and it is valid without signature.

www.oxfordlabchem.com

OXFORD LAB FINE CHEM LLP

ISO 9001-2008 Certified Company

Regd Office: Unit no 12, 1st Floor,
Neminath Industrial Estate No.6,
Navghar, Vasai (East), Palghar - 410210.
Maharashtra, INDIA.

Tel: +91 250 2390032 / 2390989 / 2390990
Email: sales@oxfordlabchem.com /
info@oxfordlabchem.com
Web: www.oxfordlabchem.com



Storage and Shelf Life : The product is highly hygroscopic; keep the container tightly closed at all times and store it properly as per the conditions mentioned on the label. The declared expiry is valid only when stored as per the conditions mentioned on the label.

Note: Sterilize media immediately after reconstitution.

Disposal: To avoid the contamination or propagation of any hazardous microbes the used, unusable or modified preparation of this product must be disposed after autoclaving after completion of task.

Reference

1. Downes F. P. and Ito K., (Ed.), (2001), *Compendium of Methods for the Microbiological Examination of Foods*, 4th Ed., American Public Health Association, Washington, D.C.
2. Handford, P. M. (1974). *A new medium for the detection and enumeration of Clostridium perfringens in foods*. *Journal of Applied Bacteriology*, 37(4), 559-570.
3. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) *Manual of Clinical Microbiology*, 11th Edition. Vol. 1.
4. Salfinger Y., and Tortorello M.L., (2015), *Compendium of Methods for the Microbiological Examination of Foods*, 5th Ed., American Public Health Association, Washington, D.C.

Disclaimer:

The information contained herein in good faith but makes no representations as to its comprehensiveness or accuracy. This document is intended only as a guide to the appropriate precautionary handling of the material by a properly trained person using this product. Individuals receiving the information must exercise their independent judgment in determining its appropriateness for a particular purpose.

Oxford Lab Fine Chem LLP makes no representations or warranties, either express or implied, including without limitation any warranties of merchantability, fitness for a particular purpose with respect to the information set forth herein or the product to which the information refers. Accordingly, Oxford Lab Fine Chem LLP will not be responsible for damages resulting from use of or reliance upon this information.

This document has been produced electronically and it is valid without signature.

www.oxfordlabchem.com